
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000541
10:30 09/07/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:11
PRESSURIZING 3:18
STERILIZING 4:27
136.0 C 220kPa 4:27
136.3 C 232kPa 4:57
136.4 C 231kPa 5:27
136.3 C 230kPa 5:57
136.2 C 231kPa 6:27
136.4 C 232kPa 6:57
136.1 C 230kPa 7:27
136.4 C 232kPa 7:57
136.4 C 232kPa 7:57
VENTING 7:57
AIR DRYING 8:51
CYCLE COMPLETE 23:51

Digital Signature #
71E48C92E527FC57
