
Midmark M11 - v1.0.5
Total Cycles: 3213

Name: RITTER2
Sterilizer ID: V2313228
12/13/2023 04:40 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 168.0 F 0.0
2:00 167.9 F 2.4
4:00 185.6 F 6.4
6:00 202.9 F 11.1
8:00 219.3 F 5.7
10:00 229.1 F 10.8
12:00 230.7 F 7.9
14:00 243.1 F 12.1
16:00 245.0 F 13.0
18:00 257.3 F 19.6
20:00 268.2 F 26.5
22:00 271.6 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.2
1:00 272.7 F 29.5
2:00 272.0 F 29.6
3:00 272.4 F 29.4
4:00 272.0 F 29.7

Min 271.6 F 29.1
Max 273.0 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.0 F 0.0
5:00 224.9 F 0.0
10:00 225.4 F 0.0
15:00 228.3 F 0.0
20:00 230.3 F 0.0
25:00 230.2 F 0.0
30:00 229.2 F 0.0
Drying Complete

FILLING: 1:39
HEATING: 22:01
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:01:14
CYCLE COMPLETE

12/13/2023 05:41 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____