
Midmark M11 - v1.0.0
Total Cycles: 5401

Name: RITTER 1
Sterilizer ID: V2178198
03/07/2025 04:54 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 164.5 F 0.0
2:00 159.1 F 3.5
4:00 202.3 F 7.7
6:00 204.6 F 13.4
8:00 224.4 F 7.2
10:00 234.0 F 12.7
12:00 235.6 F 9.3
14:00 237.1 F 9.0
16:00 247.2 F 13.9
18:00 259.5 F 20.7
20:00 270.7 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.5
1:00 272.2 F 28.6
2:00 272.1 F 29.1
3:00 271.4 F 28.6
4:00 271.8 F 28.7

Min 271.2 F 28.3
Max 272.4 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.6 F 0.0
5:00 223.3 F 0.0
10:00 212.7 F 0.0
15:00 209.9 F 0.0
20:00 224.1 F 0.0
25:00 211.3 F 0.0
30:00 217.8 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 20:42
STERILIZING: 4:00
VENTING: 3:26
DRYING: 30:00
TOTAL CYCLE: 00:59:11
CYCLE COMPLETE

03/07/2025 05:53 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____