
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 004379
10:33 19/09/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:59
PRESSURIZING 2:02
STERILIZING 2:58
136.4 C 223kPa 2:58
136.3 C 222kPa 3:28
136.2 C 221kPa 3:58
136.1 C 222kPa 4:28
136.2 C 218kPa 4:58
136.4 C 223kPa 5:28
136.4 C 224kPa 5:58
136.4 C 224kPa 6:28
136.4 C 224kPa 6:28
VENTING 6:29
AIR DRYING 7:15
CYCLE COMPLETE 7:17

Digital Signature #
424B27C97028F61F
