

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 001079  
13:56 11/01/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:39  
PRESSURIZING 3:40  
STERILIZING 4:43  
136.4 C 221kPa 4:43  
136.1 C 223kPa 5:13  
136.4 C 224kPa 5:43  
136.4 C 223kPa 6:13  
136.3 C 224kPa 6:43  
136.4 C 225kPa 7:13  
136.3 C 224kPa 7:43  
136.4 C 224kPa 8:13  
136.4 C 224kPa 8:13  
VENTING 8:13  
AIR DRYING 9:08  
CYCLE COMPLETE 9:09

-----  
Digital Signature #  
B232F1E94A6D3171  
-----