

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 001067  
14:16 09/05/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:07  
PRESSURIZING 3:17  
STERILIZING 4:56  
136.0 C 221kPa 4:56  
136.2 C 223kPa 5:26  
136.4 C 224kPa 5:56  
136.4 C 224kPa 6:26  
136.4 C 224kPa 6:56  
136.4 C 224kPa 7:26  
136.2 C 223kPa 7:56  
136.2 C 224kPa 8:26  
136.2 C 224kPa 8:26  
VENTING 8:26  
AIR DRYING 9:25  
CYCLE COMPLETE 9:51  
-----

Digital Signature #  
F1226696048B737B  
-----