
Midmark M11 - v1.0.0
Total Cycles: 2179

Name: RITTER 1
Sterilizer ID: V2178198
07/26/2022 01:03 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 155.9 F 0.0
2:00 154.6 F 0.9
4:00 169.5 F 4.7
6:00 189.5 F 9.0
8:00 217.6 F 8.6
10:00 226.3 F 8.0
12:00 238.0 F 12.7
14:00 235.6 F 9.5
16:00 240.5 F 9.2
18:00 246.8 F 13.8
20:00 258.6 F 20.3
22:00 269.3 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.7 F 29.3
2:00 271.7 F 28.4
3:00 272.5 F 29.3
4:00 272.3 F 28.9

Min 271.4 F 28.4
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.9 F 0.0
5:00 219.4 F 0.0
10:00 216.4 F 0.0
15:00 222.1 F 0.0
20:00 232.8 F 0.0
25:00 230.6 F 0.0
30:00 216.4 F 0.0
Drying Complete

FILLING: 1:15
HEATING: 23:05
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:01:33
CYCLE COMPLETE

07/26/2022 02:04 PM
ACCEPTED: X
REJECTED:
Operator: U02