
Midmark M11 - v1.0.0
Total Cycles: 2192

Name: RITTER 1
Sterilizer ID: V2178198
08/03/2022 10:27 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 186.7 F 0.0
2:00 178.9 F 3.1
4:00 191.7 F 7.3
6:00 200.3 F 11.7
8:00 216.0 F 5.7
10:00 231.5 F 10.7
12:00 233.3 F 6.7
14:00 243.7 F 12.7
16:00 241.2 F 10.8
18:00 253.6 F 17.1
20:00 264.7 F 23.7
22:00 270.9 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 28.6
1:00 272.7 F 29.2
2:00 271.7 F 28.6
3:00 272.5 F 29.1
4:00 271.8 F 28.8

Min 271.7 F 28.5
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.4 F 0.0
5:00 232.2 F 0.0
10:00 226.8 F 0.0
15:00 230.0 F 0.0
20:00 229.7 F 0.0
25:00 229.3 F 0.0
30:00 230.0 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 22:36
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:00:38
CYCLE COMPLETE

08/03/2022 11:27 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____