
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 001664
16:45 11/12/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:52
PRESSURIZING 1:58
STERILIZING 3:00
136.4 C 225kPa 3:00
136.3 C 225kPa 3:30
136.3 C 223kPa 4:00
136.2 C 224kPa 4:30
136.4 C 226kPa 5:00
136.4 C 225kPa 5:30
136.3 C 224kPa 6:00
136.4 C 226kPa 6:30
136.4 C 226kPa 6:30
VENTING 6:30
AIR DRYING 7:22
CYCLE COMPLETE 37:23

Digital Signature #
7DA5B770F15C9912
