

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2838

Name: RITTER2  
Sterilizer ID: V2313228  
09/06/2023 01:30 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 117.1 F 0.0  
2:00 119.7 F 0.0  
4:00 141.4 F 0.5  
6:00 174.0 F 4.3  
8:00 191.5 F 8.8  
10:00 210.2 F 12.3  
12:00 225.0 F 8.0  
14:00 235.4 F 13.1  
16:00 238.3 F 10.6  
18:00 235.8 F 9.0  
20:00 250.2 F 15.7  
22:00 262.0 F 22.5  
24:00 272.9 F 29.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.9  
1:00 272.8 F 29.4  
2:00 271.3 F 29.0  
3:00 272.6 F 29.4  
4:00 271.2 F 28.8

Min 271.1 F 28.7  
Max 272.9 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 182.8 F 0.0  
5:00 232.1 F 0.0  
10:00 227.2 F 0.0  
15:00 189.3 F 0.0  
20:00 184.4 F 0.0  
25:00 181.6 F 0.0  
30:00 180.8 F 0.0  
Drying Complete

FILLING: 0:55  
HEATING: 24:23  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 01:02:34  
CYCLE COMPLETE

09/06/2023 02:32 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_