
Midmark M11 - v1.0.0
Total Cycles: 4764

Name: RITTER 1
Sterilizer ID: V2178198
07/25/2024 08:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.2 F 0.0
2:00 70.9 F 0.0
4:00 83.7 F 0.1
6:00 117.7 F 1.5
8:00 143.4 F 3.7
10:00 163.3 F 6.2
12:00 177.3 F 9.3
14:00 192.2 F 12.5
16:00 207.9 F 4.0
18:00 217.9 F 7.3
20:00 228.4 F 11.0
22:00 229.5 F 6.9
24:00 238.9 F 11.1
26:00 234.9 F 8.4
28:00 248.1 F 14.4
30:00 258.6 F 20.1
32:00 268.3 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 272.3 F 29.2
2:00 272.2 F 28.8
3:00 271.4 F 28.3
4:00 271.5 F 28.7

Min 271.2 F 28.2
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.9 F 0.0
5:00 189.3 F 0.0
10:00 193.4 F 0.0
15:00 191.3 F 0.0
20:00 212.2 F 0.0
25:00 223.6 F 0.0
30:00 234.6 F 0.0

Drying Complete

FILLING: 0:47
HEATING: 33:25
STERILIZING: 4:00
VENTING: 3:48
DRYING: 30:00
TOTAL CYCLE: 01:11:60
CYCLE COMPLETE

07/25/2024 10:09 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____