
Midmark M11 - v1.0.5
Total Cycles: 3651

Name: RITTER2
Sterilizer ID: V2313228
05/28/2024 10:22 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 111.3 F 0.0
2:00 112.0 F 0.2
4:00 118.3 F 1.9
6:00 145.9 F 4.7
8:00 171.7 F 7.9
10:00 189.8 F 11.5
12:00 217.7 F 4.1
14:00 221.7 F 8.0
16:00 234.2 F 12.0
18:00 232.7 F 8.4
20:00 243.6 F 13.4
22:00 241.8 F 11.8
24:00 252.9 F 17.4
26:00 262.8 F 23.3
28:00 272.0 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.7
1:00 271.9 F 29.6
2:00 272.4 F 29.7
3:00 272.4 F 29.9
4:00 272.1 F 30.2

Min 271.5 F 29.6
Max 272.8 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.3 F 0.0
5:00 238.3 F 0.0
10:00 229.0 F 0.0
15:00 186.3 F 0.0
20:00 184.4 F 0.0
25:00 197.8 F 0.0
30:00 205.2 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 29:24
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 01:08:03
CYCLE COMPLETE

05/28/2024 11:30 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____