
Midmark M11 - v1.0.0
Total Cycles: 1402

Name: RITTER 1
Sterilizer ID: V2178198
11/29/2021 02:55 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.1 F 0.0
2:00 173.3 F 3.2
4:00 195.1 F 7.0
6:00 213.4 F 12.0
8:00 222.8 F 6.1
10:00 235.2 F 11.1
12:00 233.1 F 7.8
14:00 243.5 F 12.0
16:00 244.4 F 12.5
18:00 256.9 F 19.0
20:00 268.1 F 26.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.7 F 29.0
2:00 271.5 F 28.2
3:00 272.3 F 29.0
4:00 272.0 F 28.7

Min 271.2 F 28.1
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.3 F 0.0
5:00 227.1 F 0.0
10:00 218.2 F 0.0
15:00 209.8 F 0.0
20:00 217.2 F 0.0
25:00 216.9 F 0.0
30:00 225.7 F 0.0
Drying Complete

FILLING: 1:46
HEATING: 21:15
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:00:11
CYCLE COMPLETE

11/29/2021 04:02 PM
ACCEPTED: X
REJECTED:
Operator: U02