
Midmark M11 - v1.0.0
Total Cycles: 3521

Name: RITTER 1
Sterilizer ID: V2178198
07/13/2023 08:13 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.6 F 0.0
2:00 75.5 F 0.0
4:00 86.0 F 0.0
6:00 109.3 F 0.0
8:00 140.3 F 1.4
10:00 168.4 F 3.5
12:00 185.1 F 6.6
14:00 202.0 F 9.9
16:00 215.5 F 13.4
18:00 221.6 F 6.4
20:00 231.8 F 10.3
22:00 234.8 F 9.1
24:00 241.2 F 11.6
26:00 237.8 F 9.5
28:00 251.0 F 15.8
30:00 261.9 F 22.0
32:00 272.0 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.3
1:00 272.8 F 29.1
2:00 271.7 F 28.3
3:00 271.9 F 28.8
4:00 272.4 F 28.9

Min 271.2 F 28.1
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.3 F 0.0
5:00 201.6 F 0.0
10:00 208.3 F 0.0
15:00 217.2 F 0.0
20:00 221.6 F 0.0
25:00 222.1 F 0.0
30:00 234.6 F 0.0

Drying Complete

FILLING: 0:47
HEATING: 32:36
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:10:33
CYCLE COMPLETE

07/13/2023 09:23 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____