
Midmark M11 - v1.0.5
Total Cycles: 2353

Name: RITTER2
Sterilizer ID: V2313228
05/11/2023 09:06 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 121.3 F 0.0
2:00 125.3 F 0.0
4:00 130.9 F 0.3
6:00 167.1 F 4.1
8:00 189.2 F 8.1
10:00 214.5 F 12.6
12:00 220.9 F 7.6
14:00 233.7 F 12.5
16:00 234.2 F 9.6
18:00 237.1 F 9.4
20:00 246.5 F 14.0
22:00 258.4 F 20.5
24:00 269.1 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.1
1:00 271.7 F 29.0
2:00 272.1 F 29.4
3:00 272.2 F 29.7
4:00 271.4 F 29.4

Min 271.3 F 29.0
Max 272.5 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.3 F 0.0
5:00 205.7 F 0.0
10:00 217.9 F 0.0
15:00 225.0 F 0.0
20:00 221.8 F 0.0
25:00 219.8 F 0.0
30:00 219.6 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 24:59
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:03:07
CYCLE COMPLETE

05/11/2023 10:09 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____