
Midmark M11 - v1.0.0
Total Cycles: 3754

Name: RITTER 1
Sterilizer ID: V2178198
09/11/2023 05:29 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 148.3 F 0.0
2:00 151.2 F 0.3
4:00 183.7 F 3.8
6:00 211.7 F 8.6
8:00 227.3 F 8.8
10:00 231.7 F 9.2
12:00 238.5 F 10.3
14:00 242.7 F 12.2
16:00 240.9 F 10.8
18:00 254.7 F 17.7
20:00 266.6 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.1
1:00 272.5 F 28.7
2:00 271.9 F 28.6
3:00 271.9 F 28.2
4:00 272.7 F 28.9

Min 271.4 F 28.0
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.3 F 0.0
5:00 202.8 F 0.0
10:00 203.3 F 0.0
15:00 204.4 F 0.0
20:00 214.9 F 0.0
25:00 221.5 F 0.0
30:00 229.5 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 21:21
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:59:40
CYCLE COMPLETE

09/11/2023 06:28 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____