
Midmark M11 - v1.0.0
Total Cycles: 1980

Name: RITTER 1
Sterilizer ID: V2178198
05/31/2022 11:30 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.1 F 0.0
2:00 188.8 F 3.8
4:00 195.1 F 8.5
6:00 225.2 F 8.7
8:00 230.6 F 9.3
10:00 237.6 F 8.7
12:00 240.5 F 11.8
14:00 238.6 F 10.0
16:00 252.2 F 16.8
18:00 263.9 F 23.6
20:00 271.6 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.7
1:00 272.6 F 29.6
2:00 272.2 F 29.0
3:00 271.5 F 28.8
4:00 272.7 F 29.6

Min 271.4 F 28.6
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.3 F 0.0
5:00 212.7 F 0.0
10:00 220.8 F 0.0
15:00 225.6 F 0.0
20:00 227.9 F 0.0
25:00 228.3 F 0.0
30:00 229.0 F 0.0
Drying Complete

FILLING: 1:42
HEATING: 20:06
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:58:60
CYCLE COMPLETE

05/31/2022 12:28 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____