
Midmark M11 - v1.0.0
Total Cycles: 658

Name: MIDMARK
Sterilizer ID: V2178198
02/10/2021 08:29 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.4 F 0.0
2:00 77.9 F 0.0
4:00 85.1 F 0.0
6:00 119.3 F 0.1
8:00 163.6 F 0.2
10:00 191.6 F 1.7
12:00 214.6 F 5.3
14:00 227.0 F 8.5
16:00 236.3 F 12.5
18:00 235.0 F 8.0
20:00 244.3 F 13.5
22:00 242.7 F 11.2
24:00 239.3 F 9.6
26:00 254.1 F 16.9
28:00 266.2 F 24.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 27.7
1:00 272.1 F 28.0
2:00 272.1 F 28.5
3:00 271.6 F 27.7
4:00 272.6 F 28.5

Min 271.1 F 27.6
Max 272.7 F 28.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.4 F 0.0
5:00 220.5 F 0.0
10:00 220.4 F 0.0
15:00 216.8 F 0.0
20:00 213.0 F 0.0
Dry Cycle Aborted

FILLING: 0:50
HEATING: 29:33
STERILIZING: 4:00
VENTING: 2:55
DRYING: 22:41
TOTAL CYCLE: 00:59:59

02/10/2021 09:29 AM
ACCEPTED: X
REJECTED:
Operator: U01