

-----  
Mississauga, ON  
Canada  
-----

STATIM 5000 S501R751  
SN 511020J00004  
Unit # : 000  
WATER QUALITY  
3.3uS / 2.1ppm  
CYCLE NUMBER 000775  
13:26 16/02/2022  
WRAPPED  
134 C for 6 min.  
CYCLE START 0:00  
CONDITIONING 3:48  
PRESSURIZING 12:32  
STERILIZING 13:43  
136.1 C 222kPa 13:43  
135.9 C 221kPa 14:13  
136.3 C 225kPa 14:43  
136.4 C 225kPa 15:13  
136.4 C 223kPa 15:43  
136.4 C 227kPa 16:13  
136.4 C 224kPa 16:43  
135.8 C 222kPa 17:13  
136.4 C 225kPa 17:43  
136.0 C 222kPa 18:13  
136.4 C 227kPa 18:43  
136.2 C 222kPa 19:13  
136.3 C 225kPa 19:43  
136.3 C 225kPa 19:43  
VENTING 19:43  
AIR DRYING 20:38  
CYCLE COMPLETE 40:10

-----  
Digital Signature #  
B43B73226CF765F5  
-----