
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 007848
9:26 10/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:43
PRESSURIZING 3:44
STERILIZING 4:36
136.0 C 220kPa 4:36
136.3 C 223kPa 5:06
136.1 C 223kPa 5:36
136.2 C 221kPa 6:06
136.4 C 223kPa 6:36
136.3 C 223kPa 7:06
136.4 C 224kPa 7:36
136.3 C 221kPa 8:06
136.3 C 221kPa 8:06
VENTING 8:06
AIR DRYING 8:58
CYCLE COMPLETE 9:15

Digital Signature #
9E7943290669D4F9
