

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3177

Name: RITTER2  
Sterilizer ID: V2313228  
12/01/2023 03:34 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 95.4 F 0.0  
2:00 96.7 F 0.1  
4:00 112.0 F 1.1  
6:00 137.0 F 3.6  
8:00 166.9 F 6.9  
10:00 186.0 F 10.7  
12:00 208.2 F 5.7  
14:00 218.3 F 7.6  
16:00 229.5 F 12.0  
18:00 230.9 F 8.5  
20:00 242.1 F 12.5  
22:00 242.5 F 12.4  
24:00 254.3 F 18.3  
26:00 264.9 F 24.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.3  
1:00 272.5 F 29.8  
2:00 271.3 F 29.3  
3:00 272.7 F 30.0  
4:00 271.6 F 29.3

Min 271.2 F 29.2  
Max 272.8 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.9 F 0.0  
5:00 200.8 F 0.0  
10:00 201.7 F 0.0  
15:00 197.5 F 0.0  
20:00 192.2 F 0.0  
25:00 187.9 F 0.0  
30:00 185.4 F 0.0  
Drying Complete

FILLING: 0:58  
HEATING: 27:57  
STERILIZING: 4:00  
VENTING: 3:32  
DRYING: 30:00  
TOTAL CYCLE: 01:06:27  
CYCLE COMPLETE

12/01/2023 04:40 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_