

-----

Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 005361  
7:30 18/04/2023

UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:24  
PRESSURIZING 2:26  
STERILIZING 3:29  
136.0 C 220kPa 3:29  
136.2 C 223kPa 3:59  
136.3 C 223kPa 4:29  
136.4 C 222kPa 4:59  
136.4 C 224kPa 5:29  
136.4 C 222kPa 5:59  
136.2 C 220kPa 6:29  
136.1 C 220kPa 6:59  
136.1 C 220kPa 6:59  
VENTING 6:59  
AIR DRYING 7:46  
CYCLE COMPLETE 8:01  
-----

Digital Signature #  
EB3FCB86E8CC2CE4  
-----