

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3169

Name: RITTER2  
Sterilizer ID: V2313228  
11/29/2023 04:16 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 108.8 F 0.0  
2:00 109.9 F 0.4  
4:00 134.8 F 2.1  
6:00 164.5 F 5.5  
8:00 190.4 F 9.3  
10:00 209.5 F 10.0  
12:00 220.7 F 7.6  
14:00 231.5 F 12.1  
16:00 231.2 F 8.8  
18:00 238.9 F 11.2  
20:00 243.6 F 12.9  
22:00 255.6 F 19.2  
24:00 266.4 F 26.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.4  
1:00 272.2 F 29.8  
2:00 271.6 F 29.8  
3:00 272.1 F 29.7  
4:00 271.6 F 29.9

Min 271.2 F 29.3  
Max 272.5 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 204.2 F 0.0  
5:00 208.8 F 0.0  
10:00 194.4 F 0.0  
15:00 190.1 F 0.0  
20:00 186.9 F 0.0  
25:00 185.1 F 0.0  
30:00 183.4 F 0.0  
Drying Complete

FILLING: 1:02  
HEATING: 25:34  
STERILIZING: 4:00  
VENTING: 3:30  
DRYING: 30:00  
TOTAL CYCLE: 01:04:06  
CYCLE COMPLETE

11/29/2023 05:20 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_