

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 004971  
9:25 23/01/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:51  
PRESSURIZING 2:55  
STERILIZING 4:06  
136.4 C 221kPa 4:06  
136.2 C 222kPa 4:37  
136.3 C 222kPa 5:07  
136.3 C 222kPa 5:37  
136.3 C 222kPa 6:07  
136.0 C 219kPa 6:37  
136.3 C 223kPa 7:07  
136.4 C 221kPa 7:37  
136.4 C 221kPa 7:37  
VENTING 7:37  
CYCLE ABORTED 7:43  
STOP BUTTON PRESSED  
NOT STERILE  
END TIME 7:43  
-----

Digital Signature #  
DD45C70B1428F13B  
-----