
Midmark M11 - v1.0.0
Total Cycles: 1493

Name: RITTER 1
Sterilizer ID: V2178198
12/30/2021 11:10 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 91.7 F 0.0
2:00 122.8 F 0.0
4:00 161.8 F 0.1
6:00 178.6 F 1.3
8:00 170.7 F 3.6
10:00 202.7 F 6.3
12:00 224.1 F 10.4
14:00 221.9 F 6.6
16:00 235.7 F 9.0
18:00 244.5 F 13.6
20:00 240.9 F 10.6
22:00 235.0 F 7.9
24:00 249.3 F 14.5
26:00 260.7 F 20.9
28:00 271.2 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.3 F 28.3
2:00 272.8 F 28.8
3:00 271.6 F 28.5
4:00 272.1 F 28.3

Min 271.4 F 28.1
Max 272.8 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.4 F 0.0
5:00 225.9 F 0.0
10:00 180.1 F 0.0
Dry Cycle Aborted

FILLING: 0:55
HEATING: 29:21
STERILIZING: 4:00
VENTING: 3:07
DRYING: 11:13
TOTAL CYCLE: 00:48:36

12/30/2021 11:59 AM
ACCEPTED: X
REJECTED:
Operator: U02