

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4746

Name: RITTER 1  
Sterilizer ID: V2178198  
07/18/2024 03:23 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 107.2 F 0.0  
2:00 109.7 F 1.2  
4:00 129.4 F 4.0  
6:00 162.8 F 7.8  
8:00 196.8 F 12.7  
10:00 209.8 F 5.1  
12:00 225.2 F 9.7  
14:00 231.3 F 8.1  
16:00 238.7 F 11.2  
18:00 235.8 F 8.8  
20:00 250.2 F 15.6  
22:00 262.2 F 22.4  
24:00 272.6 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.4  
1:00 272.3 F 28.8  
2:00 272.1 F 29.0  
3:00 271.9 F 28.5  
4:00 272.6 F 29.2

Min 271.3 F 28.3  
Max 272.7 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 191.3 F 0.0  
5:00 193.5 F 0.0  
10:00 194.6 F 0.0  
15:00 194.4 F 0.0  
20:00 196.5 F 0.0  
25:00 228.0 F 0.0  
30:00 233.2 F 0.0  
Drying Complete

FILLING: 0:52  
HEATING: 24:16  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 01:02:24  
CYCLE COMPLETE

07/18/2024 04:25 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_