
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000494
13:05 21/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:50
PRESSURIZING 3:50
STERILIZING 4:54
136.4 C 222kPa 4:54
136.3 C 224kPa 5:24
136.1 C 223kPa 5:54
136.1 C 222kPa 6:24
136.4 C 225kPa 6:54
136.4 C 225kPa 7:24
136.1 C 221kPa 7:54
136.1 C 220kPa 8:24
136.1 C 220kPa 8:24
VENTING 8:24
AIR DRYING 9:16
CYCLE COMPLETE 9:28

Digital Signature #
6B565171A7DB6BA2
