

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.5uS / 0.3ppm  
CYCLE NUMBER 000430  
15:05 27/01/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:28  
PRESSURIZING 2:33  
STERILIZING 3:43  
136.5 C 233kPa 3:43  
136.4 C 233kPa 4:13  
136.3 C 233kPa 4:43  
136.4 C 232kPa 5:13  
136.2 C 232kPa 5:43  
136.4 C 232kPa 6:13  
136.4 C 233kPa 6:43  
136.3 C 231kPa 7:13  
136.3 C 231kPa 7:13  
VENTING 7:13  
AIR DRYING 8:06  
CYCLE COMPLETE 8:24  
-----

Digital Signature #  
25E8BB52AE53AEA4  
-----