
Midmark M11 - v1.0.5
Total Cycles: 1359

Name: RITTER2
Sterilizer ID: V2313228
07/27/2022 03:48 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.9 F 0.0
2:00 176.9 F 0.2
4:00 199.3 F 3.3
6:00 219.8 F 8.3
8:00 235.7 F 12.1
10:00 235.7 F 10.3
12:00 235.1 F 8.4
14:00 240.7 F 10.8
16:00 247.7 F 14.4
18:00 260.0 F 21.3
20:00 271.2 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.8
1:00 272.5 F 29.4
2:00 271.4 F 28.7
3:00 272.5 F 29.6
4:00 272.1 F 29.1

Min 271.2 F 28.7
Max 272.6 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.0 F 0.0
5:00 235.1 F 0.0
10:00 225.5 F 0.0
15:00 226.1 F 0.0
20:00 223.2 F 0.0
25:00 220.3 F 0.0
30:00 228.8 F 0.0
Drying Complete

FILLING: 1:13
HEATING: 20:43
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:59:08
CYCLE COMPLETE

07/27/2022 04:47 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____