
Midmark M11 - v1.0.5
Total Cycles: 1342

Name: RITTER2
Sterilizer ID: V2313228
07/25/2022 05:03 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 173.3 F 0.0
2:00 177.1 F 0.3
4:00 197.8 F 4.8
6:00 216.9 F 10.6
8:00 226.1 F 6.9
10:00 240.5 F 13.0
12:00 241.8 F 11.8
14:00 241.4 F 11.2
16:00 255.4 F 18.4
18:00 267.3 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.5 F 29.0
2:00 272.0 F 29.1
3:00 272.4 F 28.9
4:00 271.8 F 28.9

Min 271.5 F 28.5
Max 272.8 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.7 F 0.0
5:00 208.5 F 0.0
10:00 224.6 F 0.0
15:00 212.5 F 0.0
20:00 218.6 F 0.0
25:00 222.0 F 0.0
30:00 221.3 F 0.0
Drying Complete

FILLING: 1:25
HEATING: 19:20
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 00:57:50
CYCLE COMPLETE

07/25/2022 06:00 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____