
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002065
9:05 20/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:50
PRESSURIZING 1:57
STERILIZING 2:48
136.0 C 221kPa 2:48
136.2 C 223kPa 3:18
136.2 C 223kPa 3:48
136.1 C 223kPa 4:18
136.4 C 225kPa 4:48
136.4 C 227kPa 5:18
136.4 C 226kPa 5:48
136.3 C 225kPa 6:18
136.3 C 225kPa 6:18
VENTING 6:18
AIR DRYING 7:10
CYCLE COMPLETE 10:50

Digital Signature #
022251690CAAD1DD
