
Midmark M11 - v1.0.5
Total Cycles: 61

Name: RITTER2
Sterilizer ID: V2313228
06/01/2021 07:43 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.3 F 0.0
2:00 74.6 F 0.2
4:00 76.9 F 1.0
6:00 96.5 F 2.9
8:00 127.6 F 5.6
10:00 159.4 F 8.9
12:00 181.5 F 12.6
14:00 199.6 F 5.2
16:00 210.5 F 9.0
18:00 224.5 F 13.2
20:00 231.5 F 9.4
22:00 237.0 F 10.6
24:00 243.2 F 12.7
26:00 254.7 F 18.5
28:00 264.9 F 24.6
30:00 271.1 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 30.0
1:00 272.5 F 30.2
2:00 271.8 F 30.1
3:00 272.1 F 30.3
4:00 272.1 F 29.9

Min 271.6 F 29.9
Max 272.5 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.9 F 0.0
5:00 216.5 F 0.0
10:00 217.6 F 0.0
15:00 202.0 F 0.0
20:00 197.1 F 0.0
25:00 195.4 F 0.0
30:00 214.5 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 30:38
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:09:01
CYCLE COMPLETE

06/01/2021 08:51 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____