
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 006263
8:15 08/09/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:08
PRESSURIZING 3:12
STERILIZING 4:16
136.0 C 220kPa 4:16
136.4 C 224kPa 4:46
136.4 C 221kPa 5:16
136.4 C 221kPa 5:46
136.1 C 220kPa 6:16
136.4 C 223kPa 6:46
136.3 C 221kPa 7:16
136.2 C 217kPa 7:46
136.2 C 217kPa 7:46
VENTING 7:46
AIR DRYING 8:33
CASSETTE REMOVED
CYCLE COMPLETE 20:07

Digital Signature #
9DD0000FD5B5B48E
