

-----  
Surrey, BC  
Canada  
-----

STATIM 5000 S501R711  
SN 511018F00076  
Unit # : 000  
WATER QUALITY  
1.7uS / 1.0ppm  
CYCLE NUMBER 000024  
7:08 14/01/2020  
WRAPPED  
134 C for 6 min.  
CYCLE START 0:00  
CONDITIONING 2:34  
PRESSURIZING 14:15  
STERILIZING 15:54  
136.0 C 223kPa 15:54  
136.4 C 227kPa 16:24  
136.5 C 228kPa 16:54  
136.4 C 227kPa 17:24  
136.2 C 227kPa 17:54  
136.4 C 225kPa 18:24  
136.5 C 227kPa 18:54  
136.5 C 227kPa 19:24  
136.1 C 225kPa 19:54  
136.5 C 227kPa 20:24  
136.4 C 227kPa 20:54  
136.5 C 227kPa 21:24  
136.4 C 227kPa 21:54  
VENTING 21:54  
AIR DRYING 23:11  
CYCLE COMPLETE 53:11  
-----

Digital Signature #  
95A842EBEDE2009E  
-----