
Midmark M11 - v1.0.5
Total Cycles: 1950

Name: RITTER2
Sterilizer ID: V2313228
01/19/2023 12:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.6 F 0.0
2:00 173.5 F 3.6
4:00 186.5 F 8.0
6:00 203.4 F 13.4
8:00 223.8 F 8.5
10:00 234.1 F 10.7
12:00 239.4 F 11.7
14:00 239.3 F 10.5
16:00 252.8 F 17.4
18:00 264.6 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.2
1:00 272.6 F 29.9
2:00 271.7 F 29.2
3:00 271.5 F 29.6
4:00 273.1 F 30.4

Min 271.0 F 29.0
Max 273.1 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.2 F 0.0
5:00 227.9 F 0.0
10:00 233.2 F 0.0
15:00 225.9 F 0.0
20:00 225.9 F 0.0
25:00 225.6 F 0.0
30:00 225.1 F 0.0
Drying Complete

FILLING: 1:28
HEATING: 19:52
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:58:32
CYCLE COMPLETE

01/19/2023 01:03 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____