
Midmark M11 - v1.0.0
Total Cycles: 2571

Name: RITTER 1
Sterilizer ID: V2178198
11/16/2022 05:19 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 163.6 F 0.0
2:00 167.1 F 3.8
4:00 188.9 F 8.0
6:00 223.7 F 12.7
8:00 216.3 F 6.5
10:00 230.3 F 11.5
12:00 232.5 F 8.0
14:00 243.1 F 12.3
16:00 243.8 F 12.3
18:00 256.2 F 18.7
20:00 267.2 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.5
1:00 272.1 F 29.3
2:00 272.4 F 29.0
3:00 271.5 F 28.5
4:00 271.2 F 28.6

Min 271.1 F 28.3
Max 272.5 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.9 F 0.0
5:00 226.7 F 0.0
10:00 230.6 F 0.0
15:00 225.2 F 0.0
20:00 232.7 F 0.0
25:00 229.3 F 0.0
30:00 228.8 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 21:32
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 00:59:56
CYCLE COMPLETE

11/16/2022 06:18 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____