

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1768

Name: RITTER2  
Sterilizer ID: V2313228  
11/23/2022 12:01 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 201.0 F 0.0  
2:00 193.0 F 3.1  
4:00 210.6 F 8.8  
6:00 226.7 F 9.0  
8:00 235.6 F 11.6  
10:00 237.3 F 10.0  
12:00 237.7 F 9.2  
14:00 251.8 F 16.8  
16:00 264.4 F 24.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.0  
1:00 272.7 F 29.5  
2:00 271.6 F 29.1  
3:00 272.8 F 29.6  
4:00 271.8 F 29.0  
  
Min 271.5 F 28.9  
Max 273.0 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.1 F 0.0  
5:00 225.7 F 0.0  
10:00 230.6 F 0.0  
15:00 227.7 F 0.0  
20:00 226.9 F 0.0  
25:00 225.2 F 0.0  
30:00 223.1 F 0.0  
Drying Complete

FILLING: 2:12  
HEATING: 17:51  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:57:12  
CYCLE COMPLETE

11/23/2022 12:58 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_