

-----  
Midmark M11 - v1.0.5  
Total Cycles: 331

Name: RITTER2  
Sterilizer ID: V2313228  
09/02/2021 10:37 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 164.9 F 0.0  
2:00 166.0 F 3.3  
4:00 180.3 F 7.6  
6:00 200.0 F 12.6  
8:00 218.1 F 6.9  
10:00 231.5 F 12.1  
12:00 234.2 F 9.6  
14:00 237.9 F 9.5  
16:00 248.1 F 15.0  
18:00 259.9 F 21.7  
20:00 270.7 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.3  
1:00 272.7 F 29.8  
2:00 271.7 F 29.4  
3:00 272.9 F 29.8  
4:00 271.7 F 29.2  
  
Min 271.6 F 29.2  
Max 272.9 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 195.5 F 0.0  
5:00 202.7 F 0.0  
10:00 216.6 F 0.0  
15:00 219.1 F 0.0  
20:00 214.9 F 0.0  
25:00 216.8 F 0.0  
30:00 203.2 F 0.0  
Drying Complete

FILLING: 2:06  
HEATING: 20:46  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 01:00:02  
CYCLE COMPLETE

09/02/2021 11:37 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02