

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002550  
13:00 28/02/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:03  
PRESSURIZING 2:05  
STERILIZING 3:02  
136.5 C 224kPa 3:02  
136.3 C 226kPa 3:32  
136.4 C 226kPa 4:02  
136.2 C 223kPa 4:32  
136.3 C 224kPa 5:02  
136.4 C 225kPa 5:32  
136.4 C 224kPa 6:02  
136.4 C 226kPa 6:32  
136.4 C 226kPa 6:32  
VENTING 6:32  
AIR DRYING 7:22  
CYCLE COMPLETE 37:22  
-----

Digital Signature #  
529D7B7B3FEDCFDC  
-----