
Midmark M11 - v1.0.0
Total Cycles: 2898

Name: RITTER 1
Sterilizer ID: V2178198
02/14/2023 12:12 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 199.4 F 0.0
2:00 195.8 F 3.7
4:00 195.3 F 8.3
6:00 216.2 F 11.8
8:00 228.3 F 8.3
10:00 239.8 F 13.5
12:00 239.1 F 10.6
14:00 237.0 F 8.6
16:00 250.2 F 15.3
18:00 261.5 F 21.7
20:00 272.0 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.7 F 29.0
2:00 271.9 F 28.4
3:00 271.6 F 28.5
4:00 272.6 F 29.0

Min 271.4 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.5 F 0.0
5:00 227.3 F 0.0
10:00 231.4 F 0.0
15:00 229.1 F 0.0
20:00 230.5 F 0.0
25:00 227.5 F 0.0
30:00 227.1 F 0.0
Drying Complete

FILLING: 1:39
HEATING: 20:35
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:59:31
CYCLE COMPLETE

02/14/2023 01:11 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____