
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
1.0uS / 0.6ppm
CYCLE NUMBER 001587
8:42 11/17/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 5:31
PRESSURIZING 7:48
STERILIZING 9:34
136.0 C 220kPa 9:34
136.4 C 217kPa 10:04
136.4 C 221kPa 10:34
136.4 C 220kPa 11:05
136.3 C 221kPa 11:35
135.9 C 219kPa 12:05
136.4 C 220kPa 12:35
136.5 C 222kPa 13:04
VENTING 13:04
AIR DRYING 14:13
CYCLE COMPLETE 19:01

Digital Signature #
A4D55E1E7B136F2E
