
Midmark M11 - v1.0.0
Total Cycles: 4922

Name: RITTER 1
Sterilizer ID: V2178198
09/19/2024 12:34 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 124.5 F 0.0
2:00 124.6 F 0.1
4:00 146.3 F 0.1
6:00 173.3 F 0.7
8:00 194.4 F 3.8
10:00 208.1 F 7.3
12:00 216.3 F 11.1
14:00 224.9 F 5.6
16:00 235.2 F 10.4
18:00 234.9 F 8.1
20:00 244.6 F 12.9
22:00 242.4 F 11.4
24:00 254.4 F 17.5
26:00 264.9 F 23.7
28:00 271.3 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.2
1:00 272.4 F 28.9
2:00 272.4 F 28.6
3:00 271.4 F 28.1
4:00 271.6 F 28.5

Min 271.2 F 28.0
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.4 F 0.0
5:00 223.0 F 0.0
10:00 219.4 F 0.0
15:00 187.3 F 0.0
20:00 185.3 F 0.0
25:00 187.6 F 0.0
30:00 193.3 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 28:03
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:06:35
CYCLE COMPLETE

09/19/2024 01:40 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____