
Midmark M11 - v1.0.5
Total Cycles: 1189

Name: RITTER2
Sterilizer ID: V2313228
06/15/2022 09:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.1 F 0.0
2:00 179.2 F 3.2
4:00 183.7 F 7.2
6:00 206.9 F 12.1
8:00 220.9 F 6.9
10:00 235.1 F 12.1
12:00 235.6 F 9.7
14:00 237.1 F 9.3
16:00 248.3 F 14.9
18:00 260.3 F 21.4
20:00 271.0 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.7
1:00 272.1 F 29.2
2:00 271.2 F 28.7
3:00 273.0 F 29.8
4:00 271.9 F 29.1

Min 270.9 F 28.5
Max 273.0 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.3 F 0.0
5:00 231.9 F 0.0
10:00 233.2 F 0.0
15:00 226.4 F 0.0
20:00 223.6 F 0.0
25:00 224.0 F 0.0
30:00 222.5 F 0.0
Drying Complete

FILLING: 1:40
HEATING: 20:43
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:59:29
CYCLE COMPLETE

06/15/2022 10:39 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____