
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 004438
13:46 26/09/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:59
PRESSURIZING 3:00
STERILIZING 3:58
136.4 C 222kPa 3:58
136.4 C 226kPa 4:28
136.4 C 223kPa 4:58
136.4 C 223kPa 5:28
136.4 C 224kPa 5:58
136.1 C 223kPa 6:28
136.1 C 224kPa 6:58
136.4 C 225kPa 7:28
136.4 C 225kPa 7:28
VENTING 7:29
AIR DRYING 8:22
CYCLE COMPLETE 9:32

Digital Signature #
EB9997FAF29A64B4
