
Midmark M11 - v1.0.0
Total Cycles: 4115

Name: RITTER 1
Sterilizer ID: V2178198
12/14/2023 09:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.8 F 0.0
2:00 70.2 F 0.2
4:00 79.0 F 1.0
6:00 112.9 F 3.2
8:00 145.9 F 5.9
10:00 171.0 F 9.1
12:00 190.6 F 12.6
14:00 209.7 F 2.9
16:00 218.3 F 6.0
18:00 227.4 F 9.2
20:00 235.2 F 12.9
22:00 234.3 F 8.1
24:00 243.7 F 12.8
26:00 240.6 F 10.3
28:00 252.4 F 16.1
30:00 262.5 F 21.9
32:00 271.9 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.1
1:00 272.5 F 28.3
2:00 272.5 F 28.8
3:00 271.5 F 28.1
4:00 272.6 F 28.5

Min 271.4 F 28.0
Max 272.8 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.5 F 0.0
5:00 224.7 F 0.0
10:00 218.5 F 0.0
15:00 207.4 F 0.0
20:00 220.6 F 0.0
25:00 229.5 F 0.0
30:00 228.1 F 0.0

Drying Complete

FILLING: 0:53
HEATING: 33:16
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:11:31
CYCLE COMPLETE

12/14/2023 11:09 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____