

-----  
Midmark M11 - v1.0.5  
Total Cycles: 909

Name: RITTER2  
Sterilizer ID: V2313228  
03/25/2022 01:34 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 167.0 F 0.0  
2:00 163.7 F 0.5  
4:00 206.7 F 4.3  
6:00 223.7 F 9.5  
8:00 229.0 F 8.2  
10:00 238.0 F 11.6  
12:00 236.4 F 9.9  
14:00 236.6 F 9.0  
16:00 250.2 F 16.0  
18:00 262.6 F 23.0  
20:00 272.5 F 29.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.4  
1:00 272.8 F 29.5  
2:00 272.9 F 30.0  
3:00 271.5 F 29.4  
4:00 273.0 F 29.8  
  
Min 271.5 F 29.3  
Max 273.1 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.9 F 0.0  
5:00 232.4 F 0.0  
10:00 231.8 F 0.0  
15:00 172.7 F 0.0  
20:00 159.4 F 0.0  
25:00 164.1 F 0.0  
30:00 170.6 F 0.0  
Drying Complete

FILLING: 1:33  
HEATING: 20:54  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:59:40  
CYCLE COMPLETE

03/25/2022 02:34 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02