

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2979

Name: RITTER 1  
Sterilizer ID: V2178198  
03/08/2023 11:54 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 171.4 F 0.0  
2:00 166.0 F 3.1  
4:00 180.5 F 7.0  
6:00 201.3 F 11.9  
8:00 218.2 F 6.1  
10:00 231.6 F 11.2  
12:00 231.8 F 7.8  
14:00 242.9 F 12.1  
16:00 244.1 F 12.3  
18:00 256.7 F 18.9  
20:00 267.9 F 25.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.4  
1:00 272.3 F 28.8  
2:00 272.0 F 28.9  
3:00 272.0 F 28.6  
4:00 272.0 F 28.9

Min 271.4 F 28.3  
Max 272.9 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.2 F 0.0  
5:00 227.7 F 0.0  
10:00 235.5 F 0.0  
15:00 238.8 F 0.0  
20:00 227.3 F 0.0  
25:00 219.0 F 0.0  
30:00 216.3 F 0.0  
Drying Complete

FILLING: 0:59  
HEATING: 21:17  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 00:59:24  
CYCLE COMPLETE

03/08/2023 12:53 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_