

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 006277  
11:32 11/09/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:00  
PRESSURIZING 3:02  
STERILIZING 4:12  
136.5 C 224kPa 4:12  
136.4 C 223kPa 4:42  
136.3 C 224kPa 5:12  
136.3 C 223kPa 5:42  
136.3 C 224kPa 6:12  
136.0 C 223kPa 6:42  
136.3 C 224kPa 7:12  
136.2 C 224kPa 7:42  
136.2 C 224kPa 7:42  
VENTING 7:43  
AIR DRYING 8:34  
CASSETTE REMOVED  
CYCLE COMPLETE 26:53  
-----

Digital Signature #  
7285BE8AE6BB16D9  
-----