
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 000575
17:17 09/08/2021
WRAPPED
135 C for 10 min.
CYCLE START 0:00
CONDITIONING 1:48
PRESSURIZING 4:06
STERILIZING 5:11
136.4 C 233kPa 5:11
136.3 C 233kPa 5:41
136.4 C 233kPa 6:11
136.2 C 231kPa 6:41
136.2 C 232kPa 7:11
136.3 C 233kPa 7:41
136.2 C 232kPa 8:11
136.3 C 232kPa 8:41
136.4 C 233kPa 9:11
136.3 C 232kPa 9:41
136.4 C 233kPa 10:11
136.4 C 231kPa 10:41
136.1 C 233kPa 11:11
136.3 C 233kPa 11:41
136.4 C 230kPa 12:11
136.4 C 230kPa 12:41
136.2 C 232kPa 13:11
136.1 C 231kPa 13:41
136.4 C 232kPa 14:11
136.3 C 233kPa 14:41
136.2 C 232kPa 15:11
136.2 C 232kPa 15:11
VENTING 15:11
AIR DRYING 16:01
CYCLE COMPLETE 31:01

Digital Signature #
78E0F7A2A531FE10
