
Midmark M11 - v1.0.0
Total Cycles: 4827

Name: RITTER 1
Sterilizer ID: V2178198
08/16/2024 02:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 107.9 F 0.0
2:00 110.6 F 0.3
4:00 137.4 F 0.9
6:00 167.4 F 1.8
8:00 191.3 F 5.2
10:00 207.2 F 8.9
12:00 217.0 F 13.2
14:00 224.1 F 7.5
16:00 234.3 F 12.3
18:00 233.9 F 8.9
20:00 240.4 F 10.4
22:00 244.9 F 13.0
24:00 256.9 F 19.3
26:00 267.7 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.5
1:00 272.4 F 29.3
2:00 271.8 F 28.7
3:00 271.5 F 28.8
4:00 272.4 F 29.2

Min 271.2 F 28.4
Max 272.5 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.7 F 0.0
5:00 207.3 F 0.0
10:00 212.7 F 0.0
15:00 213.3 F 0.0
20:00 215.8 F 0.0
25:00 218.5 F 0.0
30:00 220.9 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 27:25
STERILIZING: 4:00
VENTING: 3:33
DRYING: 30:00
TOTAL CYCLE: 01:05:46
CYCLE COMPLETE

08/16/2024 03:05 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____